

Application/Control Number: 09/509,377  
 Enclosure / to Applicant's reply of Sept. 1, 2004

Page 22

Reply under 37 CFR 1.116 -  
 EXPEDITED PROCEDURE -  
 Technology Center 3739

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/509,377	08/28/2000	Sergey Matasov		9553

United States Patent and Trademark Office  
 Commissioner for Patents  
 Art Unit 3739  
 Examiner Mr. Leubecker, John P.  
 P.O. Box 1450, Alexandria VA 22313-1450  
 United States of America

EXAMINER	
LEUBECKER, JOHN P	
ART UNIT	PAPER NUMBER

3739

DATE MAILED: September 3, 2003

AMENDMENTS

Sir: In response to the Office Action of June 9, 2003, please amend the above-identified application as follows:

In the specification:

1. Please, replace the paragraph, beginning at page 3, line 12 with the following rewritten paragraph:

- A compact hollow cylinder of the invaginator can be formed of tightly compressed in longitudinal and transverse directions pleats of different forms of an eversible thin-walled tube placed at any angles with the longitudinal axis of an endoscopic tube. The cylinder has recurrent narrowings of an external diameter and widenings of its internal diameter. The stability of diameters depends on the compactness of the cylinder. In one of the embodiments the definite compactness of cylinder ensures the gap with endoscopic tube during their joining and in the process of invagination, in the other - only during the joining. There are possible also the interim variants of embodiments. -

In the claims:

Please, replace the claims 1-20 with the following amended claims:

## I claim:

1. (Amended) An endoscope, comprising an invaginator, which is a thin-walled tube, compactly placed on the distal part of an endoscopic tube in the shape of small layers and/or pleats.
2. (Amended) The endoscope according to claim 1, wherein said invaginator is formed in the shape of a compact hollow cylinder, which has a gap with the distal part of the endoscopic tube.
3. (Amended) The endoscope according to claim 2, wherein said cylinder has a compactness, which ensures said gap in the process of invagination of the endoscopic tube.
4. (Amended) The endoscope according to any of claims 1 to 3, further comprising a seal between the endoscopic tube and the unverted end of said invaginator.
5. (Amended) The endoscope according to any of claims 1 to 3, further comprising a shell of said invaginator, commensurate to the diameter of said invaginator and to the length of return.
6. (Amended) The endoscope according to any of claims 1 to 3, further comprising a preservative of the distal part of the endoscopic tube.

BEST AVAILABLE COPY

**This Page is Inserted by IFW Indexing and Scanning  
Operations and is not part of the Official Record**

**BEST AVAILABLE IMAGES**

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

- ☐ **BLACK BORDERS**
- ☐ **IMAGE CUT OFF AT TOP, BOTTOM OR SIDES**
- ☐ **FADED TEXT OR DRAWING**
- ☐ **BLURRED OR ILLEGIBLE TEXT OR DRAWING**
- ☐ **SKEWED/SLANTED IMAGES**
- ☐ **COLOR OR BLACK AND WHITE PHOTOGRAPHS**
- ☐ **GRAY SCALE DOCUMENTS**
- ☒ **LINES OR MARKS ON ORIGINAL DOCUMENT**
- ☐ **REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY**
- ☐ **OTHER:** \_\_\_\_\_

**IMAGES ARE BEST AVAILABLE COPY.**

**As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.**